



# ULTRACLEAN-3

**Alkaline concentrate for the cleaning of surgical and dental instruments. Suitable for use in ultrasonic cleaning devices.**



## Intended Purpose



ULTRACLEAN-3 is a concentrate for the cleaning of surgical and dental instruments in an ultrasonic cleaning device.

## Intended User



The product is intended for use by dental and healthcare professionals who have knowledge of the disinfection of medical devices. The product must not be used directly on patients.

## Instructions for Use



For your protection, wear a laboratory gown, household gloves and safety goggles. Prepare the working solution according to the dosing table. Run an ultrasonic cleaning cycle for 30 minutes. After cleaning, rinse the instruments thoroughly under running water of at least drinking-water quality for 10 seconds. Replace the working solution daily or when visibly contaminated. Adhere to the reprocessing recommendations of the instrument manufacturer. Not suitable for the cleaning of color-coded instruments or instruments made of natural bristles, nylon, or aluminum. A list of compatible and incompatible materials is available on our website. If you are unsure about the material composition of the medical device to be disinfected, contact the manufacturer before use. Any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

## Limitation



FOR PROFESSIONAL USE ONLY. Do not mix with other products.

## Handling



DANGER H290 May be corrosive to metals. H314 Causes severe skin burns and eye damage. P260 Do not breathe fume/mist/vapours/spray. P280 Wear protective gloves/protective clothing/eye protection/face protection. P301 + P330 + P331 IF SWALLOWED: Rinse mouth. Do NOT induce vomiting. P303 + P361 + P353 IF ON SKIN (or hair): Take off immediately all contaminated clothing. Rinse skin with water/shower. P305 + P351 + P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. P501 Dispose of contents/container in accordance with local and national regulations. Further information can be found in the Safety Data Sheet available on [oroclean.com](http://oroclean.com).



## Storage



Store in a cool, dry and well-ventilated place. Keep away from direct sunlight, heat, and open flame. Storage temperature: < 30 °C. Dispose of the product after expiry date.

## Physical Properties



Physical state: Clear, non-viscous liquid	pH-value: 14.0
Colour: Light yellow	Shelf life: 60 months (unopened container)
Scent: Aromatic	

## Composition in 100 g



100 g contains 16 g potassium hydroxide, corrosion inhibitors, auxiliaries and water.

## Dosing Table



	Concentrate	+	Water	Time
5%	50 ml	+	950 ml	30 min

\*Limited spectrum virucidal

## Compatibility



Acrylic glass	Compatible
Acrylonitrile butadiene styrene	Compatible
Chlorinated polyvinyl chloride	Compatible
Polybutylene terephthalate	Compatible
Polycarbonate	Compatible
Polyethylene	Compatible
Polyethylene terephthalate	Compatible
Polyphenylsulfone	Compatible
Polyphenylene oxide	Compatible
Polypropylene	Compatible
Polystyrene	Compatible
Polysulfone	Compatible
Polyvinyl chloride	Compatible
Polyvinylidene fluoride	Compatible
Rhodium	Compatible
Silicone	Compatible
Thermoplastic elastomer	Compatible
Thermoplastic polyurethane	Compatible
Ultra-high-molecular-weight polyethylene	Compatible



## Delivery Forms



CH-061050

1 litre bottle

## Legal Manufacturer



United Disinfectant Manufacturers AG • Dr. Grass-Strasse 12 • 9490 Vaduz • Principality of Liechtenstein  
📞 +423 237 15 03 • 📩 info@udm.li • 🌐 www.udm.li

A subsidiary of Oro Clean Chemie AG • Allmendstrasse 21 • 8320 Fehraltorf • Switzerland

📞 +41 (0)44 226 44 44 • 📩 info@oroclean.com • 🌐 www.oroclean.com

Made in Switzerland